JW

Notice of Allowability	Application No.	Applicant(s)	Applicant(s)	
	10/033,614	CHIN ET AL.		
	Examiner	Art Unit		
	Gary Jackson	3731		
The MAILING DATE of this communication appeal All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI	(OR REMAINS) CLOSED if or other appropriate commits in the commits of the commits	n this application. If not included unication will be mailed in due co	l ourse. THIS	
1. $igstyle$ This communication is responsive to <u>the "Response to the</u>	Election Requirement" filed	d April 29, 2004.		
2. X The allowed claim(s) is/are 1-6.				
3. The drawings filed on are accepted by the Examine	r.			
4.				
Attachment(s) 1. ☐ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☑ Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date 7 statements 4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. ☐ Interview S Paper No. 98), 7. ☐ Examiner's			
		Gary Jackson Primary Examiner Art Unit: 3731		

Art Unit: 3731

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows: This application is in condition for allowance except for the presence of claims 7-43 to an invention non-elected without traverse. Accordingly, claims 7-43 have been cancelled.

REASONS FOR ALLOWANCE

The following is an examiner's statement of reasons for allowance: The prior art of record fail to teach or suggest a method steps of forming a proximal anastomosis forming an aperture in a wall of the vessel; introducing through the aperture and into the vessel a flexible sealing element having a central stem and having a continuous disassociation region; expanding the sealing element to peripheral dimensions greater than the dimension of the aperture; performing anastomosis on the graft vessel; disassembling the expanded sealing element along the disassociation region.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Gary Jackson whose telephone number is (703) 308-4302. The examiner can normally be reached on Mon.-Thurs. 7:30 am to 6:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, McDermott or Shaver can be reached on (703) 308-0858. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Gary Jackson
Primary Examiner

Art Unit 3731

gj June 28, 2004